Experience gained from the RPC's credentialing programs for advanced technology clinical trials



Geoffrey S. Ibbott, Ph.D. David S. Followill, Ph.D. Andrea Molineu, MS Paola Alvarez, MS Nadia Hernandez,

Credentialing Techniques

Phantoms



QuickTime™ and a TIFF (Uncompressed) decompressor are needed to see this picture.

Benchmarks

13-15 Nov 2006



Purpose of Credentialing

- Education
- Evaluate ability to deliver dose
- Improve understanding of protocol
 - Reduce deviation rate



General Credentialing Process

- Previous patients treated with technique
- Facility Questionnaire
- Knowledge Assessment Questionnaire
- Benchmark case or phantom irradiation
- Electronic data submission
- RPC QA & dosimetry review
- Clinical review by radiation oncologist

Feedback to Institution



QuickTime[™] and a TIFF (Uncompressed) decompressor are needed to see this picture. Credentialing 3D Conformal Radiation Therapy (3D CRT)

- Evaluate 3D treatment planning process and ability to provide documentation
- 449 credentials issued to 414 distinct institutions to date
 - Most through NSABP/RTOG partial breast irradiation (PBI) trial

Credentialing LDR and HDR Brachytherapy

IAEA/QANTRM

Evaluate

- Implant technique
- Dosimetry
- Documentation
- Protocol compliance





13-15 Nov 2006

MRREAM

Credentials Awarded (based on benchmarks)

	<u>Credentials</u>	In stitutions
Prostate LDR (0232)	70	63
Prostate HDR (0321)	11	7
Breast 3D CRT (0413)	792	364
Breast Mammosite®	497	245
Breast Multicatheter	115	41
Other 3D CRT (NCCTG)	52	52
Cervix (GOG)	55	46
TOTAL	11,592	611
13-15 Nov 2006	IAEA/QANTRM	



Results of Credentialing

(closed studies)

Study	Major Deviations	Minor Deviations	Number of Patients
GOG 165 HDR Cervix Gredentialed inst	0	15	70
RTOG 95-17 HDR & LDR Breast (all)	0	4	100
RTOG 0019 LDR Prostate (values for dose only)	0	6	117 reviewed (total 129 eligible)
13-15 Nov 2006	IAEA/QAN	ΓRM	CREC MA

Results of Credentialing (closed studies)

	Major	Minor	Number of
Study	Deviations	Deviations	Patients
GOG 165 HDR Cervix			
Credentialed inst	0	15	70
Non-credentialed	57	87	275
RTOG 95-17 HDR & LDR Breast (all)	0	4	100
RTOG 0019 LDR Prostate (values for dose only)	0	6	117 reviewed (total 129 eligible)
13-15 Nov 2006	IAEA/QAN	ΓRM	CARB CAA



prostate IMRT: 4, incl. prosthesis

RPC Phantoms



thorax SBRT: 3 phantoms, 6 under construction





H&N IMRT: 20 in service, 5 under constr.

SRS: 2 in service, others sent by RDS

liver SBRT: 3, incl. motion

Scan, plan, and treat phantom as if a patient



Plan vs. Treatment



Number of Phantom Mailings



Phantom Results

Phantom	H&N	Prostate	Thorax	Liver
Irradiations	254	73	30	б
Pass	179*	55	17	3
Fail	71	9	7	1
Under analysis or at institution	30	б	6	1
Year introduced	2001	Spring 2004	Spring 2004	Spring 2005

* 30% of institutions failed H&N phantom on the first attempt

13-15 Nov 2006



Examples of Failures





13-15 Nov 2006



Explanations for Failures

Explanation	Minimum # of occurrences	
incorrect output factors in TPS	1	
incorrect PDD in TPS	1	
inadequacies in beam modeling at leaf ends (Cadman, et al; PMB 2002)	14	
not adjusting MU to account for dose differences measured with ion chamber	3	
errors in couch indexing with Peacock system	2	
2 mm tolerence on MLC leaf position	1	
setup errors	7	ADAN
target malfunction	1	

AAA

Physicists per machine





Lung Phantom Irradiations

TPS	Dose Calc. Algor correction on	Number of irradiations	Dhetero/Dhomo
Precise v 2.01	Scatter Integ. Clarkson Type	2	1.19 ± 2.6%
BrainLab	Clarkson & Pencil Beam	4	1.21 ± 0%
Eclipse	Pencil Beam	2	1.19 ± 4.6%
Ergo	3D Convolution Pencil Beam	1	1.19 ± 0.1%
Pinnacle v 6.2, 6.4 7.0g, 7.4f	Adaptative Convolve	8	1.13 ± 2.1%
Render plan	Change in primary attenuation	1	1.20
	Superposition/ Convolution	3	1.12 ± 2.4%
13-15 Nov 2006	Total IAEA/QAN	21	1.18

TLD Dose vs. Hetero Corrected Plan

TPS	Dose Calc. Algor correction on	Number of irradiations	D TLD/ D hetero
Precise v 2.01	Scatter Integ. Clarkson Type	2	0.99 ± 3.1%
BrainLab	Clarkson & Pencil Beam	4	0.96 ± 2.7%
Eclipse	Pencil Beam	2	0.97 ± 1.6%
Ergo	3D Convolution Pencil Beam	1	0.98 ± 3.2%
Pinnacle v 6.2, 6.4, 7.0g, 7.4f	Adaptative Convolve	8	0.99 ± 2.3%
Render plan	Change in primary attenuation	1	0.92
	Superposition/ Convolution	3	0.96*
	Total	21	0.97
13-15 Nov 2006	IAEA/QAI	NTRM	CARBON

The End

